

Tool ID: 1605  
Tool Location: 201

Equipment Information Sheet  
Vibro SV-10 Viscometer

**Manager:**  
**Backup:**

Xinwei Wu  
Michael Skvarla

607-254-4934  
607-254-4674

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

No safety precautions needed.

USAGE RESTRICTIONS

It is very important to measure the temperature of the fluid correctly because the viscosity is dependent upon the temperature of the fluid. The SV-10 will detect accurate temperature immediately because the fluid and the detection unit (sensor plates) with small surface area/thermal capacity reach the thermal equilibrium in a few seconds.

SCHEDULING/SIGN-UP RESTRICTIONS

None.

Minimum Tool Time: 0 minutes

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds	
Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compoundsare materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

None.

Last Updated: 07/15/2019