

Tool ID: 1611
Tool Location: 228

Equipment Information Sheet
Dektak 6M profilometer

Manager: Xinwei Wu
Backup: Jeremy Clark

607-254-4934
607-254-6487

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

- No unusual hazards during normal operation.

USAGE RESTRICTIONS
SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 15 minutes

- none

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds	
Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compoundsare materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

- Hard substrates (glass, plastic, semiconductor materials) allowed
- Soft substrates (polymers like PDMS) possibly allowed AFTER contacting tool manager.
- Sample size: up to 150 mm substrates; up to 5 mm thick

Last Updated: 08/06/2019