

Tool ID: 421
Tool Location: 133

Equipment Information Sheet
ABM Contact Aligner

Manager: Garry Bordonaro
Backup: Giovanni Sartorello

607-254-4936
607-254-4853

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

- Orange UV glasses are recommended during exposure, but not required

USAGE RESTRICTIONS

- No buddy system restrictions imposed on normal operation

SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 15 minutes

- Maximum 4 hour block reservations anytime
- Maximum 12 hours reserved in advance
- No consecutive research group reservations
- Users/Groups may use any amount of unreserved time
- Additional individual restrictions may be imposed to maximize use

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds	
Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compoundsare materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

- 4" - 9" masks, up to 200mm wafers, up to 0.25" thick
- Back of substrate must be CLEAN - NO RESIST on back

Last Updated: 08/08/2023