**Equipment Information Sheet** 

# Suss Backside Alignment Measurement

Manager: Garry Bordonaro 607-254-4936 Calls to staff phones will be automatically forwarded to

their cell phones during accessible hours. At other times leave a message or send them an email.

# SAFETY **USAGE RESTRICTIONS SCHEDULING/SIGN-UP RESTRICTIONS**

Minimum Tool Time: 0 minutes

# MATERIALS COMPATIBILITY CATEGORY

#### **Tool Category 5: Class A and B Metals and Compounds**

Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared- w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

# High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

### **Additional Material Restrictions and Exceptions**

- Adhesives are not allowed (scotch tape, Kapton tape, glues,
- epoxies, greases, etc are not allowed) in the chamber. Only use clips to secure substrates.

Last Updated: 05/07/2019