

Tool ID: 822
Tool Location: 137

Equipment Information Sheet
Suss Backside Alignment Measurement

Manager: Garry Bordonaro

607-254-4936

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

USAGE RESTRICTIONS

SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 0 minutes

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds

Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above $1e-6$ Torr at 400 C.

Additional Material Restrictions and Exceptions

- Adhesives are not allowed (scotch tape, Kapton tape, glues,
- epoxies, greases, etc are not allowed) in the chamber. Only use clips to secure substrates.

Last Updated: 05/07/2019