

Tool ID: 415
Tool Location: 121

Equipment Information Sheet

Class 1 Photolithography Spinners

Manager: Giovanni Sartorello 607-254-4853

Backup: Garry Bordonaro 607-254-4936

Backup: James Crawford 607-254-5895

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

- Face shields must be worn at ALL TIMES.
- Close spinner lid before spinning, for added protection from vacuum or substrate failure.
- No solvents or materials containing solvents outside hoods, except for immediate baking.
- Dispose of waste in appropriate receptacles.

USAGE RESTRICTIONS

- General photolithography training (online and in-person) is required to use this tool.
- Select chuck so that the **substrate is larger than the chuck's vacuum surface**.
- Inspect and clean the backside of every substrate **after coating and before baking**.
- Remove and clean chuck immediately after use.
- Clean bowl, lid and hood after use. Do not spray solvents directly onto spindle.

SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 0 minutes

- None

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds	
Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

- Class 1 materials **only** (most positive resist, nLOF, i-line and DUV materials, P-20 primer).
- NO WATER.
- No biomaterials.