

Tool ID: 419
Tool Location: 133

Equipment Information Sheet

SUSS MA6-BA6 Contact Aligner

Manager:

Garry Bordonaro
Michael Skvarla

607-254-4936
607-254-4674

~~Calls to staff phones will be automatically forwarded to~~
their cell phones during accessible hours. At other times
leave a message or send them an email.

SAFETY

- Orange UV glasses are recommended during exposure, but not required.

USAGE RESTRICTIONS

- No buddy system restriction imposed on normal operation.

SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 30 minutes

- Maximum 4 hour block reservations anytime
- No consecutive research group reservations
- Additional individual restriction may be imposed to maximize use

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds

Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared- w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

- 5" & 7" mask size, up to 150mm substrates size.
- Backside of the substrate must be CLEAN - No Resist on the back.

Last Updated: 11/26/2024