

Tool ID: 624
Tool Location: 111 West Wall

Equipment Information Sheet
Woollam RC2 Spectroscopic Ellipsometer

Manager: Alan R. Bleier
Backup: Jeremy Clark
Backup: Tom Pennell

607-254-4931
607-254-6487
607-254-4309

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

- Do not stare into the light beam, eye injury could result.
- Put a white piece of paper in the beam to locate the beam.

USAGE RESTRICTIONS

- No buddy system restrictions imposed on normal operation.

SCHEDULING/SIGN-UP RESTRICTIONS

- Maximum 4 hour reservation - anytime

Minimum Tool Time: 10 minutes

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds

Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

Some tool restrictions on high vapor pressure materials may apply.

Last Updated: 03/23/2022