

Tool ID: 101
Tool Location: 103

Equipment Information Sheet
General Material Anneal 1 - A1

Manager:

Phil Infante
Backup Windsor

607-254-4926
607-254-4831

~~Calls to staff phones will be automatically forwarded to~~
their cell phones during accessible hours. At other times
leave a message or send them an email.

SAFETY

- The furnaces are run at elevated temperatures of 400-1200°C and use flammable, toxic, and corrosive gases.

USAGE RESTRICTIONS

- No changing of gas flows or process parameters without staff approval
- Max temperature of 1100 C

SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 90 minutes

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds

Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

Last Updated: 03/20/2019