

Tool ID: 320
Tool Location: 114

Equipment Information Sheet

Angstrom E-Beam Evaporator

Manager: Aaron Windsor 607-254-4831

Backup: Christopher Alpha 607-254-4913

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

USAGE RESTRICTIONS

- Must remain in cleanroom when electron beam is on.

SCHEDULING/SIGN-UP RESTRICTIONS

MATERIALS COMPATIBILITY CATEGORY

Minimum Tool Time: 30 minutes

Tool Category 3: Silcion, III-V Compound Semiconductor, Glass and Metal Category	
Allowed	Not Allowed
Tool category 1/1E and 2 materials	Glass Substrates
III/V compound Semiconductors allowed	No CNF Class A or Class B metals-and oxides/compounds of i.e. (Magnesium,Zinc, Barium, Calcium)
PECVD and ALD Films	No High Vapor pressure materials
Cured organics and baked Photoresist	
Organic/Bio Materials prepped w/o Salt Buffers	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

Last Updated: 09/06/2022