

Tool ID: 622
Tool Location: 103

Equipment Information Sheet

Keithley 4200A - IVCV Testing Station

Manager: Phil Infante 607-254-4926
Backup: Jeremy Clark 607-254-6487

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

- The CV station measurement chuck can be heated to 300°C, please be aware of the chuck temp before use.
- No other unusual hazards associated with the tool.

USAGE RESTRICTIONS

SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 12 minutes

- See Coral

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds	
Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

- None

Last Updated: 11/30/2018