

Tool ID: 600
Tool Location: 103

Equipment Information Sheet

Everbeing EB-6 DC Probe Station

Manager: Phil Infante 607-254-4926
Backup: Jeremy Clark 607-254-6487
Backup: Aaron Windsor 607-254-4831

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

- The IV station measurement chuck can be heated to 300°C, please be aware of the chuck temp before use.
- The system meters can generate 110 V and 100mA of output potential/current.
- No other unusual hazards associated with the tool.

USAGE RESTRICTIONS

SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 12 minutes

- None

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds

| Allowed | Not Allowed |
|---|---|
| Tool category 1/1E, 2, 3, and 4 materials | |
| Silicon Based Substrates and Films | |
| III/V compound Semiconductors | |
| Glass Substrates | |
| PECVD and ALD Films | |
| Cured organics and baked Photoresist | |
| CNF Class A, B, and Refractory metals | |
| Exposed Gold, Silver, Copper | |
| Alkali and Alkaline Compounds | |
| Organic/Biology Molecules prepared-w/salt buffers | |
| High Vapor Pressure Materials (Mg, Ca, Zn)* | * Some tool restrictions on high vapor pressure materials may apply |
| Soft organic materials | |

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

- None

Last Updated: 04/10/2026