

Tool ID: 600  
Tool Location: 103

Equipment Information Sheet

# Everbeing EB-6 DC Probe Station

**Manager: Phil Infante 607-254-4926**  
**Backup: Jeremy Clark 607-254-6487**  
**Backup: Aaron Windsor 607-254-4831**

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

**SAFETY**

- The IV station measurement chuck can be heated to 300°C, please be aware of the chuck temp before use.
- The system meters can generate 110 V and 100mA of output potential/current.
- No other unusual hazards associated with the tool.

**USAGE RESTRICTIONS**

**SCHEDULING/SIGN-UP RESTRICTIONS**

*Minimum Tool Time: 12 minutes*

- None

**MATERIALS COMPATIBILITY CATEGORY**

Tool Category 5: Class A and B Metals and Compounds	
Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

**High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.**

**Additional Material Restrictions and Exceptions**

- None

*Last Updated: 08/06/2019*