

Tool ID: 1603
Tool Location: 201

Equipment Information Sheet

Rame-Hart 500 Goniometer

Manager: Xinwei Wu **607-254-4934**

Backup: Jeremy Clark **607-254-6487**

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

SAFETY

- No unusual hazards during normal operation.

USAGE RESTRICTIONS

SCHEDULING/SIGN-UP RESTRICTIONS

Minimum Tool Time: 30 minutes

- None

MATERIALS COMPATIBILITY CATEGORY

Tool Category 5: Class A and B Metals and Compounds	
Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

Additional Material Restrictions and Exceptions

Last Updated: 11/29/2018