

Tool ID: 619  
Tool Location: 128

Equipment Information Sheet

# Zygo Optical Profilometer

**Manager:** Philip Schneider 607-254-4931

**Backup:** Christopher Alpha 607-254-4913

**Backup:** John Treichler 607-254-4949

Calls to staff phones will be automatically forwarded to their cell phones during accessible hours. At other times leave a message or send them an email.

**SAFETY**

- There are no hazards during the normal use of this instrument.

**USAGE RESTRICTIONS**

**SCHEDULING/SIGN-UP RESTRICTIONS**

Minimum Tool Time: 15 minutes

- Maximum 2 hour reservation between 9AM and 5PM, Monday through Friday.

**MATERIALS COMPATIBILITY CATEGORY**

Tool Category 5: Class A and B Metals and Compounds	
Allowed	Not Allowed
Tool category 1/1E, 2, 3, and 4 materials	
Silicon Based Substrates and Films	
III/V compound Semiconductors	
Glass Substrates	
PECVD and ALD Films	
Cured organics and baked Photoresist	
CNF Class A, B, and Refractory metals	
Exposed Gold, Silver, Copper	
Alkali and Alkaline Compounds	
Organic/Biology Molecules prepared-w/salt buffers	
High Vapor Pressure Materials (Mg, Ca, Zn)*	* Some tool restrictions on high vapor pressure materials may apply
Soft organic materials	

High Vapor Pressure Metals and Compounds are materials that have a vapor pressure above 1e-6 Torr at 400 C.

**Additional Material Restrictions and Exceptions**

- No material restrictions.
- Curved, sloped, and vertical edges may not scan properly.

Last Updated: 06/09/2025